



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0256
 CALIBRATION DATE: 21-Sep-21

Glider Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.922437e-001 CPcor = -9.5700e-008
 h = 1.254969e-001 CTcor = 3.2500e-006
 i = -1.346138e-004 WBOTC = -1.4808e-007
 j = 2.585577e-005


| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000 | 0.0000 | 0.00000 | 2813.81 | 0.00000 | 0.00000 |
| 1.0000 | 34.6422 | 2.96250 | 5612.33 | 2.96253 | 0.00003 |
| 4.5000 | 34.6232 | 3.26830 | 5824.45 | 3.26825 | -0.00004 |
| 15.0000 | 34.5815 | 4.24585 | 6455.32 | 4.24582 | -0.00002 |
| 18.5000 | 34.5727 | 4.58953 | 6662.68 | 4.58957 | 0.00005 |
| 24.0000 | 34.5644 | 5.14528 | 6984.57 | 5.14527 | -0.00001 |
| 29.0000 | 34.5595 | 5.66497 | 7272.39 | 5.66497 | -0.00000 |
| 32.5000 | 34.5553 | 6.03560 | 7470.79 | 6.03583 | 0.00022 |

$$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$$

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

$$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0256
 CALIBRATION DATE: 14-Sep-21

Glider Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.920914e-001 CPcor = -9.5700e-008
 h = 1.254960e-001 CTcor = 3.2500e-006
 i = -1.367531e-004 WBOTC = -1.4808e-007
 j = 2.598962e-005


| BATH TEMP (° C) | BATH SAL (PSU) | BATH COND (S/m) | INSTRUMENT OUTPUT (Hz) | INSTRUMENT COND (S/m) | RESIDUAL (S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000 | 0.0000 | 0.00000 | 2813.66 | 0.00000 | 0.00000 |
| 1.0000 | 34.6431 | 2.96257 | 5612.45 | 2.96258 | 0.00002 |
| 4.5000 | 34.6237 | 3.26834 | 5824.60 | 3.26833 | -0.00001 |
| 15.0000 | 34.5830 | 4.24601 | 6455.53 | 4.24595 | -0.00006 |
| 18.4999 | 34.5746 | 4.58974 | 6662.97 | 4.58980 | 0.00006 |
| 24.0000 | 34.5656 | 5.14544 | 6984.84 | 5.14545 | 0.00001 |
| 29.0000 | 34.5606 | 5.66513 | 7272.65 | 5.66512 | -0.00001 |
| 32.5000 | 34.5569 | 6.03585 | 7471.01 | 6.03587 | 0.00002 |

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0256
 CALIBRATION DATE: 10-Sep-21

Glider Payload CTD PRESSURE CALIBRATION DATA
 1450 psia S/N 10719843

COEFFICIENTS:

| | | | |
|-----------|----------------|---------|----------------|
| PA0 = | 2.828849e-001 | PTCA0 = | 5.245004e+005 |
| PA1 = | 4.510012e-003 | PTCA1 = | 4.357900e+000 |
| PA2 = | -1.917590e-011 | PTCA2 = | -8.430304e-002 |
| PTEMPA0 = | -6.123978e+001 | PTCB0 = | 2.515668e+001 |
| PTEMPA1 = | 5.350856e-002 | PTCB1 = | -8.478803e-004 |
| PTEMPA2 = | -6.653089e-007 | PTCB2 = | 0.000000e+000 |

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | THERMISTOR OUTPUT (volts) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) | TEMP (°C) | THERMISTOR OUTPUT (volts) | INSTRUMENT OUTPUT (counts) |
|-----------------|----------------------------|---------------------------|--------------------------|-----------------|------------------|---------------------------|----------------------------|
| 14.49 | 527711.2 | 1616.0 | 14.52 | 0.00 | 32.50 | 1792 | 527920.00 |
| 301.49 | 591291.0 | 1619.5 | 301.41 | -0.01 | 29.00 | 1724 | 527922.50 |
| 588.65 | 654983.7 | 1621.1 | 588.66 | 0.00 | 24.00 | 1626 | 527925.60 |
| 875.85 | 718702.1 | 1623.1 | 875.86 | 0.00 | 18.50 | 1519 | 527919.70 |
| 1162.99 | 782432.1 | 1624.7 | 1162.97 | -0.00 | 15.00 | 1451 | 527911.40 |
| 1450.10 | 846195.3 | 1625.8 | 1450.07 | -0.00 | 4.50 | 1248 | 527887.90 |
| 1163.04 | 782456.1 | 1626.4 | 1163.08 | 0.00 | 1.00 | 1180 | 527870.60 |
| 875.88 | 718711.9 | 1627.0 | 875.91 | 0.00 | | | |
| 588.74 | 655006.1 | 1627.3 | 588.76 | 0.00 | TEMPERATURE (°C) | SPAN | |
| 301.62 | 591323.2 | 1630.0 | 301.56 | -0.00 | -5.10 | 25.16 | |
| 14.49 | 527710.5 | 1630.6 | 14.52 | 0.00 | 35.00 | 25.13 | |

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$


$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 10-Sep-21 0.00





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0256
 CALIBRATION DATE: 21-Sep-21

Glider Payload CTD TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.827970e-004
 a1 = 3.218334e-004
 a2 = -5.497910e-006
 a3 = 2.290690e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000 | 565222.3 | 1.0000 | -0.0000 |
| 4.5000 | 483947.9 | 4.5000 | 0.0000 |
| 15.0000 | 309755.2 | 14.9999 | -0.0001 |
| 18.5000 | 268608.0 | 18.5001 | 0.0001 |
| 24.0000 | 215992.5 | 24.0000 | 0.0000 |
| 29.0000 | 178242.8 | 28.9999 | -0.0001 |
| 32.5000 | 156334.6 | 32.5000 | 0.0000 |

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0256
 CALIBRATION DATE: 14-Sep-21

Glider Payload CTD TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -1.864013e-004
 a1 = 3.223627e-004
 a2 = -5.513669e-006
 a3 = 2.287861e-007

| BATH TEMP (° C) | INSTRUMENT OUTPUT (counts) | INST TEMP (° C) | RESIDUAL (° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000 | 565250.0 | 1.0000 | 0.0000 |
| 4.5000 | 483965.2 | 4.5000 | -0.0000 |
| 15.0000 | 309761.2 | 14.9997 | -0.0003 |
| 18.4999 | 268606.4 | 18.5004 | 0.0005 |
| 24.0000 | 215996.2 | 24.0000 | -0.0000 |
| 29.0000 | 178249.6 | 28.9996 | -0.0004 |
| 32.5000 | 156339.8 | 32.5002 | 0.0002 |

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature

