



Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9309  
 CALIBRATION DATE: 19-Mar-22


Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.770825e-001      CPcor = -9.5700e-008  
 h = 1.417013e-001      CTcor = 3.2500e-006  
 i = -2.215016e-004      WBOTC = 5.7053e-007  
 j = 3.733068e-005

| BATH TEMP<br>(° C) | BATH SAL<br>(PSU) | BATH COND<br>(S/m) | INSTRUMENT<br>OUTPUT (Hz) | INSTRUMENT<br>COND (S/m) | RESIDUAL<br>(S/m) |
|--------------------|-------------------|--------------------|---------------------------|--------------------------|-------------------|
| 22.0000            | 0.0000            | 0.00000            | 2628.90                   | 0.00000                  | 0.00000           |
| 1.0000             | 34.6814           | 2.96553            | 5277.20                   | 2.96552                  | -0.00001          |
| 4.5000             | 34.6621           | 3.27161            | 5477.54                   | 3.27161                  | 0.00001           |
| 15.0000            | 34.6206           | 4.25014            | 6073.05                   | 4.25014                  | 0.00000           |
| 18.5000            | 34.6119           | 4.59417            | 6268.71                   | 4.59418                  | 0.00001           |
| 24.0000            | 34.6022           | 5.15029            | 6572.35                   | 5.15027                  | -0.00002          |
| 29.0000            | 34.5963           | 5.67033            | 6843.80                   | 5.67034                  | 0.00001           |
| 32.5000            | 34.5916           | 6.04122            | 7030.93                   | 6.04150                  | 0.00028           |

f = Instrument Output(Hz) \* sqrt(1.0 + WBOTC \* t) / 1000.0  
 t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;  
 Conductivity (S/m) = (g + h \* f<sup>2</sup> + i \* f<sup>3</sup> + j \* f<sup>4</sup>) / (1 + δ \* t + ε \* p)  
 Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9309  
 CALIBRATION DATE: 19-Mar-22

Slocum Payload CTD TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:


a0 = -1.000401e-004  
 a1 = 3.045231e-004  
 a2 = -4.255517e-006  
 a3 = 1.968585e-007

| BATH TEMP<br>(° C) | INSTRUMENT<br>OUTPUT (counts) | INST TEMP<br>(° C) | RESIDUAL<br>(° C) |
|--------------------|-------------------------------|--------------------|-------------------|
| 1.0000             | 571780.5                      | 1.0000             | 0.0000            |
| 4.5000             | 489240.4                      | 4.4999             | -0.0001           |
| 15.0000            | 312533.6                      | 15.0000            | 0.0000            |
| 18.5000            | 270846.3                      | 18.5001            | 0.0001            |
| 24.0000            | 217580.1                      | 23.9999            | -0.0001           |
| 29.0000            | 179390.8                      | 29.0000            | 0.0000            |
| 32.5000            | 157243.9                      | 32.5000            | 0.0000            |

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

$$\text{Residual (°C)} = \text{instrument temperature} - \text{bath temperature}$$





Sea-Bird Scientific  
 13431 NE 20<sup>th</sup> Street  
 Bellevue, WA 98005  
 USA

+1 425-643-9866  
 seabird@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 9309  
 CALIBRATION DATE: 16-Mar-22

Slocum Payload CTD PRESSURE CALIBRATION DATA  
 1450 psia S/N 4386337

COEFFICIENTS:

|           |                |         |                |
|-----------|----------------|---------|----------------|
| PA0 =     | 6.792636e-001  | PTCA0 = | 5.254094e+005  |
| PA1 =     | 4.618943e-003  | PTCA1 = | -5.983539e+000 |
| PA2 =     | -1.469401e-011 | PTCA2 = | 5.246470e-001  |
| PTEMPA0 = | 1.569842e+002  | PTCB0 = | 2.516213e+001  |
| PTEMPA1 = | -6.652269e-002 | PTCB1 = | -1.745636e-004 |
| PTEMPA2 = | 1.871659e-007  | PTCB2 = | 0.000000e+000  |

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

| PRESSURE (PSIA) | INSTRUMENT OUTPUT (counts) | THERMISTOR OUTPUT (volts) | COMPUTED PRESSURE (PSIA) | RESIDUAL (%FSR) | TEMP (°C) | THERMISTOR OUTPUT (volts) | INSTRUMENT OUTPUT (counts) |
|-----------------|----------------------------|---------------------------|--------------------------|-----------------|-----------|---------------------------|----------------------------|
| 14.68           | 528573.3                   | 2019.9                    | 14.62                    | -0.00           | 32.50     | 1881                      | 528770.90                  |
| 302.71          | 590958.4                   | 2016.8                    | 302.74                   | 0.00            | 29.00     | 1935                      | 528679.90                  |
| 587.86          | 652708.4                   | 2016.9                    | 587.83                   | -0.00           | 24.00     | 2010                      | 528570.40                  |
| 874.72          | 714865.4                   | 2014.8                    | 874.68                   | -0.00           | 18.50     | 2094                      | 528489.60                  |
| 1161.75         | 777092.6                   | 2012.8                    | 1161.73                  | -0.00           | 15.00     | 2147                      | 528440.50                  |
| 1448.66         | 839327.9                   | 2012.9                    | 1448.73                  | 0.00            | 4.50      | 2307                      | 528379.20                  |
| 1161.80         | 777095.0                   | 2013.5                    | 1161.75                  | -0.00           | 1.00      | 2361                      | 528419.30                  |
| 874.81          | 714885.1                   | 2015.4                    | 874.77                   | -0.00           |           |                           |                            |
| 587.86          | 652715.2                   | 2017.0                    | 587.86                   | 0.00            |           |                           |                            |
| 300.88          | 590601.0                   | 2016.9                    | 301.09                   | 0.01            |           |                           |                            |
| 14.68           | 528575.1                   | 2018.8                    | 14.62                    | -0.00           |           |                           |                            |

|  |                  |       |
|--|------------------|-------|
|  | TEMPERATURE (°C) | SPAN  |
|  | -5.00            | 25.16 |
|  | 35.10            | 25.16 |

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

● 16-Mar-22 0.00

