



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9409
 CALIBRATION DATE: 07-Mar-23


Slocum Payload CTD CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.828525e-001 CPcor = -9.5700e-008
 h = 1.211957e-001 CTcor = 3.2500e-006
 i = -1.492606e-004 WBOTC = 2.3388e-007
 j = 2.608070e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2850.24	0.00000	0.00000
1.0000	34.8111	2.97556	5715.06	2.97559	0.00003
4.5000	34.7913	3.28260	5931.74	3.28258	-0.00002
15.0000	34.7489	4.26422	6575.97	4.26420	-0.00002
18.5000	34.7392	4.60924	6787.59	4.60924	-0.00001
24.0000	34.7277	5.16690	7116.00	5.16692	0.00002
29.0000	34.7191	5.68819	7409.47	5.68823	0.00004
32.5000	34.7073	6.05913	7611.15	6.05910	-0.00003

f = Instrument Output(Hz) * sqrt(1.0 + WBOTC * t) / 1000.0
 t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;
 Conductivity (S/m) = (g + h * f² + i * f³ + j * f⁴) / (1 + δ * t + ε * p)
 Residual (Siemens/meter) = instrument conductivity - bath conductivity





Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9409
 CALIBRATION DATE: 03-Mar-23

Slocum Payload CTD PRESSURE CALIBRATION DATA
 1450 psia S/N 10230819

COEFFICIENTS:

PA0 = 1.251182e-001	PTCA0 = 5.244067e+005
PA1 = 4.564224e-003	PTCA1 = 2.640462e+000
PA2 = -1.901125e-011	PTCA2 = -6.185740e-002
PTEMPA0 = -6.867528e+001	PTCB0 = 2.506801e+001
PTEMPA1 = 5.198644e-002	PTCB1 = 8.020050e-004
PTEMPA2 = -5.010503e-007	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

THERMAL CORRECTION

PRESSURE (PSIA)	INSTRUMENT OUTPUT (counts)	THERMISTOR OUTPUT (volts)	COMPUTED PRESSURE (PSIA)	RESIDUAL (%FSR)	TEMP (°C)	THERMISTOR OUTPUT (volts)	INSTRUMENT OUTPUT (counts)
14.46	527615.0	1770.6	14.63	0.01	32.50	1984	527663.60
301.61	590535.4	1772.8	301.53	-0.01	29.00	1914	527668.60
588.80	653565.4	1773.8	588.78	-0.00	24.00	1814	527604.00
876.02	716633.2	1774.4	876.04	0.00	18.50	1705	527664.20
1163.29	779735.2	1774.6	1163.32	0.00	15.00	1635	527679.60
1450.49	842835.6	1774.8	1450.43	-0.00	4.50	1427	527666.00
1163.31	779744.2	1774.4	1163.36	0.00	1.00	1358	527609.80
876.09	716654.6	1774.0	876.14	0.00			
588.88	653591.0	1773.4	588.89	0.00			
301.70	590558.0	1773.2	301.63	-0.00			
14.46	527560.4	1772.2	14.38	-0.01			

	TEMPERATURE (°C)	SPAN
	-5.00	25.06
	34.90	25.10

y = thermistor output (counts)

$$t = PTEMPA0 + PTEMPA1 * y + PTEMPA2 * y^2$$

$$x = \text{instrument output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$


$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressure (PSIA)} = PA0 + PA1 * n + PA2 * n^2$$

$$\text{Residual (\%FSR)} = (\text{computed pressure} - \text{true pressure}) * 100 / \text{Full Scale Range}$$

Date, Offset (%FSR)

- 12-Sep-17 -0.02
- ▲ 03-Mar-23 0.00



**POST CRUISE
 CALIBRATION**



Sea-Bird Scientific
 13431 NE 20th Street
 Bellevue, WA 98005
 USA

+1 425-643-9866
 seabird@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 9409
 CALIBRATION DATE: 07-Mar-23

Slocum Payload CTD TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = -2.221646e-004
 a1 = 3.291908e-004
 a2 = -5.999896e-006
 a3 = 2.405514e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	571543.0	1.0000	0.0000
4.5000	489410.2	4.5000	-0.0000
15.0000	313350.2	15.0002	0.0002
18.5000	271766.2	18.4997	-0.0003
24.0000	218572.2	24.0002	0.0002
29.0000	180409.6	28.9999	-0.0001
32.5000	158258.6	32.5000	0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$$

$$\text{Residual (°C)} = \text{instrument temperature} - \text{bath temperature}$$

