

PO Box 518  
620 Applegate St  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## SLC Testing Certification

Date 5/8/2018

S/N# FLBBCDSL-5059

### Low temperature test #1

Chill 2.5 hr at -20 °C

### High temperature test #1 Verify operation post-testing

Heat 2.5 hr at 50 °C

### Low temperature test #2 same protocol as #1

### High temperature test #2 same protocol as #1

### Vacuum test

< 0.1" Hg change in 10 min.

### Pressure test

5 cycles, 0–1250 m with 10-sec. soaks

Held at 1250 m for 2 hrs. on last cycle

### Electrical isolation

Resistance between copper faceplate and grounding wire is > 1 mΩ

### Calibration verification

Verify calibration and dark counts in bb, chl, and CDOM channels

Verify 5% of single point check for chl and bb

Verify 10% of single point check for CDOM

Signature

*Ann Caidis-Morgan*

NOTES:

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## ECO CDOM Fluorometer Characterization Sheet

Date: 5/7/2018

S/N: FLBBCDSL-5059

CDOM concentration expressed in ppb can be derived using the equation:

$$\text{CDOM (ppb)} = \text{Scale Factor} * (\text{Output} - \text{Dark Counts})$$

|   |                  |
|---|------------------|
| Dark Counts                                 | Digital          |
| Scale Factor (SF)                           | 46 counts        |
| Maximum Output                              | 0.0905 ppb/count |
| Resolution                                  | 4130 counts      |
|   | 1.0 counts       |
| Ambient temperature during characterization | 22.5 °C          |

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**SF:** Determined using the following equation:  $SF = x / (\text{output} - \text{dark counts})$ , where  $x$  is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

**Maximum Output:** Maximum signal output the fluorometer is capable of.

**Resolution:** Standard deviation of 1 minute of collected data.

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## ECO Chlorophyll Fluorometer Characterization Sheet

Date: 5/7/1948

S/N: FLBBCDSL-5059

Chlorophyll concentration expressed in  $\mu\text{g/l}$  can be derived using the equation:

$$\text{CHL } (\mu\text{g/l}) = \text{Scale Factor} * (\text{Output} - \text{Dark counts})$$

|   |                              |
|---|------------------------------|
| Dark counts                                 | Digital                      |
| Scale Factor (SF)                           | 43 counts                    |
| Maximum Output                              | 0.0073 $\mu\text{g/l/count}$ |
| Resolution                                  | 4130 counts                  |
|   | 1.0 counts                   |
| Ambient temperature during characterization | 22.5 °C                      |

**Dark Counts:** Signal output of the meter in clean water with black tape over detector.

**SF:** Determined using the following equation:  $\text{SF} = x / (\text{output} - \text{dark counts})$ , where x is the concentration of the solution used during instrument characterization. SF is used to derive instrument output concentration from the raw signal output of the fluorometer.

**Maximum Output:** Maximum signal output the fluorometer is capable of.

**Resolution:** Standard deviation of 1 minute of collected data.

The relationship between fluorescence and chlorophyll-a concentrations in-situ is highly variable. The scale factor listed on this document was determined using a mono-culture of phytoplankton (*Thalassiosira weissflogii*). The population was assumed to be reasonably healthy and the concentration was determined by using the absorption method. To accurately determine chlorophyll concentration using a fluorometer, you must perform secondary measurements on the populations of interest. This is typically done using extraction-based measurement techniques on discrete samples. For additional information on determining chlorophyll concentration see "Standard Methods for the Examination of Water and Wastewater" part 10200 H, published jointly by the American Public Health Association, American Water Works Association, and the Water Environment Federation.

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



SEA-BIRD  
SCIENTIFIC

(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## Scattering Meter Calibration Sheet

5/7/2018

Wavelength: 700

S/N FLBBCDSL-5059

Use the following equation to obtain either digital or analog "scaled" output values:

$$\beta(\theta_c) \text{ m}^{-1} \text{ sr}^{-1} = \text{Scale Factor} \times (\text{Output} - \text{Dark Counts})$$

|                           |   |  |
|---------------------------|---|--|
| • Scale Factor for 700 nm | = | 2.090E-06 (m <sup>-1</sup> sr <sup>-1</sup> )/counts |
| • Output                  | = | meter output counts                                  |
| • Dark Counts             | = | 44 counts  |
| Instrument Resolution     | = | 1.1 counts   |

### Definitions:

- **Scale Factor:** Calibration scale factor,  $\beta(\theta_c)/\text{counts}$ . Refer to User's Guide for derivation.
  - **Output:** Measured signal output of the scattering meter.
  - **Dark Counts:** Signal obtained by covering detector with black tape and submersing sensor in water.
- Instrument Resolution: Standard deviation of 1 minute of collected data.